



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

OLYMPUS MEDICAL SYSTEMS CORP.

% Daphney Germain-Kolawole

Regulatory Affairs Project Manager

Olympus America, Inc.

3500 Corporate Parkway, P.O. Box 610

Center Valley, PA 18034-0610

JUL 27 2015

Re: K130058

Trade/Device Name: Endoscopic Ultrasound Center EU-Y0008

Regulation Number: 21 CFR 892.1550

Regulation Name: Ultrasonic pulsed doppler imaging system

Regulatory Class: II

Product Code: IYN, ITX, IYO, ODG

Dated (Date on orig SE ltr): January 8, 2013

Received (Date on orig SE ltr): January 9, 2013

Dear Daphney Germain-Kolawole,

This letter corrects our substantially equivalent letter of February 22, 2013.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be

found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Benjamin R. Fisher -S

Benjamin R. Fisher, Ph.D.
Director
Division of Reproductive, Gastro-Renal,
and Urological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use Form

510(k) Number (if known): K130058

Device Name: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Indications for Use:

This ultrasound center is intended to be used with Olympus ultrasound endoscopes, Olympus ultrasound probes or Olympus esophageal ultrasound probes to observe and to store real-time ultrasound images and indicated for use within the gastrointestinal (GI) tract, biliary and pancreatic ducts and surrounding organs, airways and tracheobronchial tree, and urinary tract.

Prescription Use ✓
(Part 21 CFR 801, Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of *In Vitro* Diagnostics and Radiological Health (OIR)

(Division Sign Off)
Division of Radiological Health
Office of *In Vitro* Diagnostic and Radiological Health
510(k) _____

Page 1 of 21

1.1.4.1 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	N		N		N(*2)	N (*3)	N (*4)
	Trans-vaginal							
	Trans-urethral	N		N		N(*2)	N (*3)	N (*4)
	Trans-esoph. (non-Card.)	N		N		N(*2)	N (*3)	N (*4)
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	N		N		N(*2)	N (*3)	N (*4)
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract, biliary, pancreatic duct and surrounding organs,
Intraluminal ultrasound for airways and tracheobronchial tree

*2: Includes Power Doppler

*3: Combination of each operating mode, B, PWD, Color Doppler and Other

*4: 3-D Imaging, Harmonic Imaging

(Division Sign Off)

Division of Radiological Health

Office of *In Vitro* Diagnostic and Radiological Health
510(k)

Section 1.1 Administrator:

Indications for Use Statement

Page 3 of 22

1.1.4.2 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC PROBE RU-75M-R1

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P						
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Card.)							
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify)							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

(Division Sign Off)
Division of Radiological Health
Office of *In Vitro* Diagnostic and Radiological Health
510(k)

1.1.4.3 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC PROBE RU-12M-R1

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P						
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Card.)							
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify)							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
Peripheral Vessel	Other (Specify)							
	Peripheral vessel							

N = new indication; P = previously cleared by FDA; E = added under this appendix

(Division Sign Off)
Division of Radiological Health
Office of *In Vitro* Diagnostic and Radiological Health
510(k)

1.1.4.4 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC PROBE UM-2R

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P						
	Trans-vaginal							
	Trans-urethral	P						
	Trans-esoph. (non-Card.)	P						
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	P						
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract, biliary, pancreatic duct and surrounding organs,
Intraluminal ultrasound for upper airways and tracheobronchial tree

(Division Sign Off)
Division of Radiological Health
Office of *In Vitro* Diagnostic and Radiological Health
510(k)

1.1.4.5 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC PROBE UM-3R

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P						
	Trans-vaginal							
	Trans-urethral	P						
	Trans-esoph. (non-Card.)	P						
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	P						
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract, biliary, pancreatic duct and surrounding organs,
Intraluminal ultrasound for upper airways and tracheobronchial tree

(Division Sign Off)

Division of Radiological Health

Office of *In Vitro* Diagnostic and Radiological Health
510(k)

Section 1.1 Administrative Information
Indications for Use Statement
Page 7 of 22

1.1.4.6 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC PROBE UM-S20-20R

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P						
	Trans-vaginal							
	Trans-urethral	P						
	Trans-esoph. (non-Card.)	P						
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	P						
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract, biliary, pancreatic duct and surrounding organs,
Intraluminal ultrasound for upper airways and tracheobronchial tree

(Division Sign Off)

Division of Radiological Health

Office of In Vitro Diagnostic and Radiological Health

510(k)

Section 1.1 Administrative Information

Indications for Use Statement

Page 8 of 22

1.1.4.7 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC PROBE UM-G20-29R

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P						
	Trans-vaginal							
	Trans-urethral	P						
	Trans-esoph. (non-Card.)	P						
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	P						
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract, biliary, pancreatic duct and surrounding organs,
Intraluminal ultrasound for upper airways and tracheobronchial tree

(Division Sign Off)

Division of Radiological Health

Office of *In Vitro* Diagnostic and Radiological Health
510(k)

Section 1.1 Administrative Information
Indications for Use Statement
Page 9 of 22

1.1.4.8 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC PROBE UM-BS20-26R

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:								
Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P						
	Trans-vaginal							
	Trans-urethral	P						
	Trans-esoph. (non-Card.)	P						
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
Intravascular								
Other (Specify) (*1)	P							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
Other (Specify)								
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract, biliary, pancreatic duct and surrounding organs,
Intraluminal ultrasound for upper airways and tracheobronchial tree

(Division Sign Off)

Division of Radiological Health

Office of *In Vitro* Diagnostic and Radiological Health

510(k)

Section 1.1 Administrative Information

Indications for Use Statement

Page 10 of 22

1.1.4.9 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC PROBE UM-S20-17S

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P						
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Card.)	P						
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	P						
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract wall ;biliary duct(common bile, cystic, intrahepatic); pancreatic ducts ;
and surrounding organs; upper airways and tracheobronchial tree.

(Division Sign Off)

Division of Radiological Health

Office of *In Vitro* Diagnostic and Radiological Health

510(k)

Section 1.1 Administrative Information

Indications for Use Statement

Page 11 of 22

1.1.4.10 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC PROBE UM-S30-20R

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P						
	Trans-vaginal							
	Trans-urethral	P						
	Trans-esoph. (non-Card.)	P						
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	P						
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract, biliary, pancreatic duct and surrounding organs,
Intraluminal ultrasound for upper airways and tracheobronchial tree

1.1.4.11 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC PROBE UM-S30-25R

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P						
	Trans-vaginal							
	Trans-urethral	P						
	Trans-esoph. (non-Card.)	P						
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	P						
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Intraluminal ultrasound for Gastrointestinal tract and surrounding organs,
upper airways and tracheobronchial tree

1.1.4.12 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC PROBE UM-DP12-25R

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P						P(*2)
	Trans-vaginal							
	Trans-urethral	P						P(*2)
	Trans-esoph. (non-Card.)	P						P(*2)
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	P						P(*2)
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract, biliary, pancreatic duct and surrounding organs,
Intraluminal ultrasound for upper airways and tracheobronchial tree

*2: 3-D Imaging

1.1.4.13 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC PROBE UM-DP20-25R

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P						P(*2)
	Trans-vaginal							
	Trans-urethral	P						P(*2)
	Trans-esoph. (non-Card.)	P						P(*2)
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
Cardiac	Intravascular							
	Other (Specify) (*1)	P						P(*2)
	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
Peripheral Vessel	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract, biliary, pancreatic duct and surrounding organs,
Intraluminal ultrasound for upper airways and tracheobronchial tree

*2: 3-D Imaging

(Division Sign Off)

Division of Radiological Health

Office of *In Vitro* Diagnostic and Radiological Health

510(k)

Section 1.1 Administrative Information

Indications for Use Statement

Page 15 of 22

1.1.4.14 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC PROBE UM-DG20-31R

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P						P(*2)
	Trans-vaginal							
	Trans-urethral	P						P(*2)
	Trans-esoph. (non-Card.)	P						P(*2)
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
Cardiac	Intravascular							
	Other (Specify) (*1)	P						P(*2)
	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
Peripheral Vessel	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract, biliary, pancreatic duct and surrounding organs

*2: 3-D Imaging

(Division Sign Off)

Division of Radiological Health

Office of *In Vitro* Diagnostic and Radiological Health

510(k)

Section 1.1 Administrative Information

Indications for Use Statement

Page 16 of 22

1.1.4.15 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: EUS EXERA ULTRASONIC GASTROVIDEOSCOPE OLYMPUS GF TYPE UM160

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Card.)	P						
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	P						
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract, biliary, pancreatic duct and surrounding organs

1.1.4.16 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: EVIS EXERA II ULTRASONIC BRONCHOFIBERVIDEOSCOPE OLYMPUS
BF TYPE UC180F

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Card.)	P		P		P(*2)	P(*3)	P(*4)
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	P		P		P(*2)	P(*3)	P(*4)
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
Peripheral Vessel	Other (Specify)							
	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Intraluminal ultrasound for airways and tracheobronchial tree

*2: Includes Power Doppler

*3: Combination of each operating mode, B, PWD, Color Doppler and Other

*4: Harmonic Imaging

(Division Sign Off)

Division of Radiological Health

1.1.4.17 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC GASTROVIDEOSCOPE OLYMPUS GF TYPE UE160-AL5

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Card.)	P		P		P(*2)	P(*3)	P(*4)
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	P		P		P(*2)	P(*3)	P(*4)
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract, biliary, pancreatic duct and surrounding organs

*2: Includes Power Doppler

*3: Combination of each operating mode, B, PWD, Color Doppler and Other

*4: Harmonic Imaging

Section 1.1 Administrative Information

Indications for Use Statement

Page 19 of 22

(Division Sign Off)

Division of Radiological Health

Office of *In Vitro* Diagnostic and Radiological Health

510(k)

1.1.4.18 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC GASTROVIDEOSCOPE OLYMPUS GF TYPE UC140P-AL5

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Card.)	P		P		P(*2)	P(*3)	
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	P		P		P(*2)	P(*3)	
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract, biliary, pancreatic duct and surrounding organs

*2: Includes Power Doppler

*3: Combination of each operating mode, B, PWD and Color Doppler

(Division Sign Off)

Division of Radiological Health

Office of *In Vitro* Diagnostic and Radiological Health

510(k)

1.1.4.19 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: ULTRASONIC GASTROVIDEOSCOPE OLYMPUS GF TYPE UCT140-AL5

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Card.)	P		P		P(*2)	P(*3)	
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	P		P		P(*2)	P(*3)	
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (Specify)							
Peripheral Vessel	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract, biliary, pancreatic duct and surrounding organs

*2: Includes Power Doppler

*3: Combination of each operating mode, B, PWD and Color Doppler

(Division Sign Off)

Division of Radiological Health

Office of *In Vitro* Diagnostic and Radiological Health

510(k)

1.1.4.20 Diagnostic Ultrasound Indications For Use Format

System: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008

Transducer: EVIS EXERA II ULTRASOUND GASTROVIDEOSCOPE OLYMPUS GF
TYPE UCT180

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intra-operative (Specify)							
	Intra-operative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (Specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Card.)	P		P		P(*2)	P(*3)	P(*4)
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
	Other (Specify) (*1)	P		P		P(*2)	P(*3)	P(*4)
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
Peripheral Vessel	Other (Specify)							
	Peripheral vessel							
	Other (Specify)							

N = new indication; P = previously cleared by FDA; E = added under this appendix

Additional Comments:

*1: Specification for "Other"

Gastrointestinal tract and surrounding organs

*2: Includes Power Doppler

*3: Combination of each operating mode, B, PWD, Color Doppler and Other

*4: Harmonic Imaging



(Division Sign Off)

Division of Radiological Health

Office of *In Vitro* Diagnostic and Radiological Health
510(k) K130058

Section 1.1 Administrative Information
Indications for Use Statement
Page 22 of 22

K130058

510(k) Summary

FEB 22 2013

1. General Information

- Applicant: OLYMPUS MEDICAL SYSTEMS CORP.
2951 Ishikawa-cho, Hachioji-shi, Tokyo,
Japan 192-8507
- Official Correspondent: Daphney Germain-Kolawole
Regulatory Affairs Project Manager
Olympus America, Inc.
3500 Corporate Parkway
PO Box 610
Center Valley, PA 18034-0610, USA
Phone: 484-896-5691
FAX: 484-896-7128
- Prepared Date: January 8, 2013

2. Device Identification

- Device Name: ENDOSCOPIC ULTRASOUND CENTER EU-Y0008
- Common Name: Diagnostic Ultrasound System
- Regulation Number: 892.1550 Ultrasonic pulsed doppler imaging system
892.1560 Ultrasonic pulsed echo imaging system
892.1570 Diagnostic ultrasound transducer
876.1500 Endoscope and Accessories
- Regulatory Class: II
- Product Code: IYN, IYO, ITX, and KOG

3. Predicate Devices

- ENDOSCOPIC ULTRASOUND CENTER EU-Y0006 (K121564)
- HI VISION Preirus Diagnostic Ultrasound Scanner (K093466)

4. Device Description

The subject system, EU-Y0008 ENDOSCOPIC ULTRASOUND CENTER, combined with Ultrasound videoscopes or Ultrasound probes to make an endoscopic ultrasound imaging system that can acquire and display high-resolution and high-penetration, real-time ultrasound images of the target organs.

The subject system has modes of B, PWD, Color Doppler, Combined and Others (3-D Imaging and Harmonic Imaging) and Elastography function which visualizes the amount of strain in tissue (hardness of tissue) during compression and retraction.

The subject system provides measurements and calculations of distance, area, circumference, volume, time and blood velocity and provides storage and retrieval of images for review and printing. The subject system enables the user to print images and record images to external recording devices. Additionally, the subject system enables the user to record movies to internal memory.

The subject system can identify and recognize compatible Olympus transducers and display endoscopic images in addition to ultrasound images.

The basic design, system configuration, general operation, and user interface of this subject system are substantially equivalent to the predicate devices.

5. Indications for Use

This ultrasound center is intended to be used with Olympus ultrasound endoscopes, Olympus ultrasound probes or Olympus esophageal ultrasound probes to observe and to store real-time ultrasound images and indicated for use within the gastrointestinal (GI) tract, biliary and pancreatic ducts and surrounding organs, airways and tracheobronchial tree, and urinary tract.

6. Technological Characteristics

The subject system, EU-Y0008 ENDOSCOPIC ULTRASOUND CENTER, is comparable with and substantially equivalent to the EU-Y0006 ENDOSCOPIC ULTRASOUND CENTER (K121564) and the Diagnostic Ultrasound Scanner HITACHI HI VISION Preirus (K093466). The key technological differences from the EU-Y0006 (K121564) in comparison to the EU-Y0008 (subject) are the addition of the Elastography function, TIC Analysis function, and movie storage. These new functions are present in the Hitachi HI VISION Preirus (K093466) device. The subject system has the same technological characteristics, key features, indications for use, and mode of operation as the predicate devices. The subject system has been designed to meet the applicable safety standards.

7. Conclusion

When compared to the predicate devices, the EU-Y0008 ENDOSCOPIC ULTRASOUND CENTER does not incorporate any significant changes in intended use, mode of operation, or design that could affect the safety or effectiveness of the device.